

論文 / 著書情報  
Article / Book Information

Title(English)	Development of a new device control system for $^1\text{H}$ -NMR experiments
Authors(English)	Naoki Yoshida, Hideki Ueno, Akihiro Yoshimi, Yoko Ishibashi, Yuichi Ichikawa, Yasushi Abe, KOICHIRO ASAH, Masatoshi Chikamori, Tomomi Fujita, Takeshi Furukawa, Eri Hikota, Yuichi Ohtomo, Yuta Saito, Hazuki Shirai, Takahiro Suzuki, Xiaofei Yang
Citation(English)	XVI International Conference on Electromagnetic Isotope Separators and Techniques Related to their Applications (EMIS 2012)
Pub. date	2012, 12
URL	<a href="http://ribf.riken.go.jp/emis2012/">http://ribf.riken.go.jp/emis2012/</a>
Copyright	Copyright (c) 2012 Elsevier

## Development of a new device control system for $\beta$ -NMR experiments

N. Yoshida<sup>a</sup>, H. Ueno<sup>b</sup>, A. Yoshimi<sup>c</sup>, Y. Ishibashi<sup>b,d</sup>, Y. Ichikawa<sup>a</sup>, Y. Abe<sup>b,d</sup>,  
K. Asahi<sup>a</sup>, M. Chikamori<sup>a</sup>, T. Fujita<sup>f</sup>, T. Furukawa<sup>e</sup>, E. Hikota<sup>a</sup>,  
D. Nagae<sup>d</sup>, Y. Ohtomo<sup>a</sup>, Y. Saito<sup>d</sup>, H. Shirai<sup>a</sup>, T. Suzuki<sup>a</sup>, X.F. Yang<sup>b</sup>

<sup>a</sup>Department of Physics, Tokyo Institute of Technology, Japan

<sup>b</sup>RIKEN Nishina Center, Japan

<sup>c</sup>Research Core for Extreme Quantum World, Okayama University, Japan

<sup>d</sup>Department of Physics, University of Tsukuba, Japan

<sup>e</sup>Department of Physics, Tokyo Metropolitan University, Japan

<sup>f</sup>Department of Physics, Osaka University, Japan

*Key Words:  $\beta$ -NMR, electromagnetic moments,*

We have newly developed a device-control system of a  $\beta$ -NMR apparatus for the measurement of nuclear electromagnetic moments of rare isotopes. In this development, we installed a fast switching system into a high-power tank circuit to apply a strong oscillating magnetic field [1] and a new Programmable Sequence Generator module for its control. The system enables to conduct an NMR frequency scan with wide frequency windows by dynamically changing the circuit's resonant frequency. In addition, logical electric-circuit modules of the data taking system were replaced by a CPLD module, in which their functions can be programmed.

The performance of the new system was studied in a  $\beta$ -NMR experiment with spin-polarized  $^{20}\text{F}$  isotopes carried out at the RIBF facility. The  $^{20}\text{F}$  isotope was produced in the  $(d, p)$  reaction with a polarized deuteron beam at  $E/A = 7$  MeV on a  $\text{CaF}_2$  target. Details of the new  $\beta$ -NMR system and the experimental result will be given in the presentation.

### References

[1] K. Minamisono *et al.*, Nucl. Instr. and Meth. in Phys. Res. A **589**, 185 (2008).