

論文 / 著書情報
Article / Book Information

Title(English)	Development of a new device control system for ^1H -NMR experiments
Authors(English)	Naoki Yoshida, Hideki Ueno, Akihiro Yoshimi, Yoko Ishibashi, Yuichi Ichikawa, Yasushi Abe, KOICHIRO ASAH, Masatoshi Chikamori, Tomomi Fujita, Takeshi Furukawa, Eri Hikota, Yuichi Ohtomo, Yuta Saito, Hazuki Shirai, Takahiro Suzuki, Xiaofei Yang
Citation(English)	XVI International Conference on Electromagnetic Isotope Separators and Techniques Related to their Applications (EMIS 2012)
Pub. date	2012, 12
URL	http://ribf.riken.go.jp/emis2012/
Copyright	Copyright (c) 2012 Elsevier

Development of a new device control system for β -NMR experiments

N. Yoshida^a, H. Ueno^b, A. Yoshimi^c, Y. Ishibashi^{b,d}, Y. Ichikawa^a, Y. Abe^{b,d},
K. Asahi^a, M. Chikamori^a, T. Fujita^f, T. Furukawa^e, E. Hikota^a,
D. Nagae^d, Y. Ohtomo^a, Y. Saito^d, H. Shirai^a, T. Suzuki^a, X.F. Yang^b

^aDepartment of Physics, Tokyo Institute of Technology, Japan

^bRIKEN Nishina Center, Japan

^cResearch Core for Extreme Quantum World, Okayama University, Japan

^dDepartment of Physics, University of Tsukuba, Japan

^eDepartment of Physics, Tokyo Metropolitan University, Japan

^fDepartment of Physics, Osaka University, Japan

Key Words: β -NMR, electromagnetic moments,

We have newly developed a device-control system of a β -NMR apparatus for the measurement of nuclear electromagnetic moments of rare isotopes. In this development, we installed a fast switching system into a high-power tank circuit to apply a strong oscillating magnetic field [1] and a new Programmable Sequence Generator module for its control. The system enables to conduct an NMR frequency scan with wide frequency windows by dynamically changing the circuit's resonant frequency. In addition, logical electric-circuit modules of the data taking system were replaced by a CPLD module, in which their functions can be programmed.

The performance of the new system was studied in a β -NMR experiment with spin-polarized ^{20}F isotopes carried out at the RIBF facility. The ^{20}F isotope was produced in the (d, p) reaction with a polarized deuteron beam at $E/A = 7$ MeV on a CaF_2 target. Details of the new β -NMR system and the experimental result will be given in the presentation.

References

[1] K. Minamisono *et al.*, Nucl. Instr. and Meth. in Phys. Res. A **589**, 185 (2008).